



Guilford County Purchasing Department Specifications - Event 387- No Substitute Bid

Section I - General

Guilford County is seeking **No Substitute Bids** to establish a Price Only Contract for: Stryker Power Pro XT 6506 Stretchers. The purpose of this bid package is, therefore, to obtain competitive bids in accordance with the North Carolina Purchasing Laws and the Guilford County Purchasing Policies.

The Price Only Contract shall be initially for two years with the option to renew for two additional, one year renewal periods. This is a Price Only Contract, which means there are no guarantees as to the minimum or maximum amount ordered and dollar amount expended.

Cancellation Clause: Either party may cancel the contract upon a 30 day written notice to cancel for no apparent reasons.

Bonds: No bid deposit or performance bond will be required for this bid process.

Pre-Bid Meeting: There will not be a pre-bid conference held in connection with this bid package since it is believed that the specifications contained herein fully cover the purchase intent.

ALL QUESTIONS must be submitted electronically by means of the Q & A Feature included within this bidding event. NO QUESTIONS will be answered by telephone, email, written correspondence or other means except as described herein. Answers to electronic questions properly submitted will be sent out electronically by means of the above described Q & A feature to all properly registered suppliers who received the Initial Notification of bidding event. Those suppliers who registered after the Event Preview date and time will not receive electronic correspondence relative to this Event. There shall be **no communication** with any Guilford County employees other than by means of the Q & A feature. Violation of this requirement may invalidate your proposal.

Terms and Conditions: Terms and Conditions, the Non-Collusion Affidavit and the E-Verify Affidavit are attached to the event and should be reviewed carefully.

ALL Bids are to be submitted online via www.co.guilford.nc.us by the event close date and time of 2:00 PM on Thursday, October 17, 2013. Also, note that any hard copy information may be attached online or hand carried to:

- Guilford County Purchasing
- Attn: Shayla Parker, MWBE Coordinator/Buyer
- Old County Courthouse, Suite B-32, 301 West Market Street
- Greensboro, North Carolina 27401



Section II – Specifications

This is a **No Substitute** bid for the purchase of Stryker Power Pro XT stretchers. These stretchers must contain the following and should be reflected in the price of the stretcher:

POWER-PRO XT 3 Stage IV Pole (Patient Right)
POWER-PRO XT Equipment Hook
POWER-PRO XT Storage Flat (Head End)
POWER-PRO XT Adjustable Wheel Lock (Dual)
POWER-PRO XT SMRT 110V AC Option
POWER-PRO XT SMRT Charger Mounting Bracket
POWER-PRO XT In-Fastener Shut-Off Bracket
POWER-PRO XT Litter (Std Trendelenburg)
POWER-PRO XT Mattress (G-Rated)
POWER-PRO XT Restraint Package (G Rated)
POWER-PRO XT Safety Hook (Short)

The Stryker Power Pro XT stretcher must have the following dimensions:

Height ¹	
Highest Position	41.5 in (105 cm)
Lowest Position	14 in (36 cm)
Length	
Standard	81 in (206 cm)
Minimum	63 in (160 cm)
Width	23 in (58 cm)
Weight ²	125 lb (57 kg)
Wheels	
Diameter	6 in (15cm)
Width	2 in (5 cm)
Articulation	
Backrest	0-73°
Shock Position	+15°
Optional Knee Gatch	30°
Maximum Weight Capacity ³	700 lb (kg)
Minimum Operator Required	
Occupied Cot	2
Unoccupied Cot	1



Recommended Fastener System

Power-LOAD™

Model 6390

Floor Mount

Model 6370 or 6377

Wall Mount

Model 6371

Recommended Loading Height ⁴

¹ Height measured from bottom of mattress, at seat section, to ground level

² Cot is to be weighed with one battery pack, without mattress and restraints.

³ 700 lb weight capacity with an unassisted life capacity of 500 lb

⁴ To accommodate load decks up to 36 in. Load height option to be set between 26 in and 36 in.

Warranty: A warranty is to be included on all parts, labor, travel, soft goods, X-frame components, powertrain and on all welds.